

Nama Alat : **Dental X-ray**  
Merk/Type : **Öwandy / Altis DC**  
Negara Asal : **France**

## **TECHNICAL SPECIFICATION**

Device	Altis DC Class II B 93/42 EN 60601-1, EN 60601-1-1 EN 60601-1-2, EN 60601-1-3 EN 60601-1-4, EN 60601-2-7 EN 60601-2-28, EN 60601-2-32 EN 1441
Voltage	220/230/240 V $\pm$ 10% / 50-60 Hz
Pre-programmed exposure time	0.01 – 2 seconds (35 increments)
Tube current	4 mA or 5 mA
Focal length	0.7 mm (IEC 336)
Distance from focal spot to skin	20 cm (standard) 30 cm (with cone extension)
Tension	High-frequency 65 kV, DC (Inverse frequency: 100 kHz)
Total filtration	> 2 mm Al eq. @ 65 kW
Timer	Powered by microchip 60 pre-programmed exposure time
Serial interface	RS 232 serial port for future applications
Extension arm	30, 60, 80 cm
Dimensions when fully extended	1,495 mm/ 1,795 mm/ 1,995 mm
Handswitch	with extension cable (3 m) or directly on the timer

The EC symbol certifies that the product complies with the 93/42/EEC directive